Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 2355
CALIBRATION DATE: 06-May-21

SBE 37 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

```
g = -1.034113e+000
h =  1.536166e-001
i = -1.718256e-004
j =  4.160420e-005

CPcor = -9.5700e-008
CTcor =  3.2500e-006
WBOTC =  7.7050e-006
```

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2595.75	0.00000	0.00000
0.9998	34.7227	2.96871	5101.20	2.96872	0.00001
4.4999	34.7028	3.27506	5291.89	3.27505	-0.00001
14.9999	34.6608	4.25454	5859.25	4.25455	0.00001
18.5000	34.6523	4.59896	6045.78	4.59894	-0.00002
23.9999	34.6432	5.15571	6335.40	5.15571	0.00001
28.9999	34.6382	5.67641	6594.38	5.67642	0.00001
32.5000	34.6354	6.04800	6772.93	6.04800	-0.00001

f = Instrument Output(Hz) * sqrt(1.0 + WBOTC * t) / 1000.0

t = temperature (°C); p = pressure (decibars); δ = CTcor; ε = CPcor;

Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \varepsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity