



CALIBRATION CERTIFICATE

Form No. 622, Dec 2005

a xylem brand

Sensing Foil Batch No: 1206
Certificate No: 3835 1033 1260

Product: 3835
Serial No: 1094
Calibration Date: July 16, 2014

This is to certify that this product has been calibrated using the following instruments:

Fluke CHUB E-4	Serial No. A7C677
Fluke 5615 PRT	Serial No. 849155
Fluke 5615 PRT	Serial No. 802054
Honeywell PPT	Serial No. 44074
Calibration Bath model FNT 321-1-40	1

Parameter: Internal Temperature:

Calibration points and readings:

Temperature (°C)	-	-	-	-
Reading (mV)	-	-	-	-

Giving these coefficients

Index	0	1	2	3
TempCoef	2.57969E+01	-3.08597E-02	2.91577E-06	-4.33313E-09

*Note: Temperature calibration NOT performed

Parameter: Oxygen:

	O2 Concentration	Air Saturation
Range:	0-500 μM ¹⁾	0 - 120%
Accuracy ¹⁾ :	< $\pm 8\mu\text{M}$ or $\pm 5\%$ (whichever is greater)	$\pm 5\%$
Resolution:	< 1 μM	< 0.4%
Settling Time (63%):	< 25 seconds	

Calibration points and readings²⁾:

	Air Saturated Water	Zero Solution (Na_2SO_3)
Phase reading (°)	3.09206E+01	6.42677E+01
Temperature reading (°C)	1.01548E+01	1.98673E+01
Air Pressure (hPa)	1.02100E+03	

Giving these coefficients

Index	0	1	2	3
PhaseCoef	-1.35301E+00	1.14268E+00	0.00000E+00	0.00000E+00

¹⁾ Valid for 0 to 2000m (6562ft) depth, salinity 33 - 37ppt

²⁾ The calibration is performed in fresh water and the salinity setting is set to: 0

Date:
July 16, 2014

Signed: Jason B Baxter

Aanderaa Data Instruments

1700/1725 Brannum Lane | Yellow Springs, OH 45387 | Tel. +1-937-767-4974 | email: infoUSA@xyleminc.com