



SEA-BIRD ELECTRONICS, INC.

13431 NE 20th St. Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

Pressure Test Certificate

Customer Pacific Marine Environmental Lab
Job Number 66476A
Date 1/27/2012
Technician JK

<u>Serial Number</u>	164208-0655
----------------------	-------------

<u>Low Pressure (PSI)</u>	50	PSI
<u>Time (Minutes)</u>	15	Minutes

<u>High Pressure (PSI)</u>	1000	PSI
<u>Time (Minutes)</u>	30	Minutes

Pass ☒

Fail ☐

Comments

Replaced 6-pin connector. Replaced the main piston "O"-Rings.

