

Certificate no:  
Oxygen Optode 4330\_4099\_207074

Product: Oxygen Optode 4330  
Serial no: 4099  
Test Date: 02.11.2022

This is to certify that this product has been pressure tested:

Reference Equipment	Serial
Autoklav 800 bar	210005

**Pressure Cycle Test:**

Pressure (kPa)	Duration (Minutes)	Number of Cycles
60000	60	1

Pressure (kPa)	Duration (Minutes)
0	0

No irregularities were found during the test

Date:02.11.2022

Erlend S. Lid

Erlend Lid, Production Engineer