



SEA-BIRD ELECTRONICS, INC.

13431 NE 20th St. Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

Pressure Test Certificate

Customer Pacific Marine Environmental Lab

Job Number 63663

Date 4/11/2011

Technician JK

Serial Number 164208-0653

Low Pressure (PSI) 50 PSI

Time (Minutes) 15 Minutes

High Pressure (PSI) 1000 PSI

Time (Minutes) 30 Minutes

Pass ☒

Fail ☐

Comments

Replaced 6-pin connector. Replaced the main piston "O"-Rings.

