

# CALIBRATION CERTIFICATE

a xylem brand

**Layout No:**  
**Range:** 0 - 20 MPa

**Product:** Pressure Sensor 3815D  
**Serial No:** 683  
**Calibration Date:** October 2, 2014

## Calibration

The calibration of each sensor is carried out at the factory and the calibration is valid for all following data handling. The form below is filled in only where necessary, depending upon which type of sensor it concerns.

Pressure		Coefficients (kPa)	
Bar	Reading N		
Air	38	A	-5.503E-01
50	327	B	1.632E-02
100	570	C	3.756E-06
150	811	D	0.000E+00
200	1023		

Date:  
October 2, 2014

Sign: Jason Baxter