Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 1526
CALIBRATION DATE: 07-Aug-22

SBE 37 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

```
g = -1.040900e+000
h =  1.596948e-001
i = -1.161906e-004
j =  3.809446e-005

CPcor = -9.5700e-008
CTcor =  3.2500e-006
WBOTC =  5.9407e-006
```

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2553.27	0.00000	0.00000
1.0000	34.6665	2.96438	5002.26	2.96440	0.00002
4.5000	34.6475	3.27036	5188.94	3.27034	-0.00002
15.0000	34.6068	4.24862	5744.48	4.24860	-0.00002
18.5000	34.5986	4.59260	5927.19	4.59259	-0.00000
24.0000	34.5897	5.14863	6210.87	5.14865	0.00002
29.0000	34.5845	5.66861	6464.57	5.66863	0.00002
32.5000	34.5804	6.03949	6639.40	6.03947	-0.00002

f = Instrument Output(Hz) * sqrt(1.0 + WBOTC * t) / 1000.0

t = temperature (°C); p = pressure (decibars); δ = CTcor; ε = CPcor;

Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \varepsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity