

CERTIFICATE of CALIBRATION for LI-COR SENSOR

Spherical Quantum Sensor
Model Number: LI-193

Serial Number: SPQA3290

Calibration Date: May 01, 2019
Manufacture Date: December 05, 2002

Calibration Constants:

Output (in-air): 7.15 microamps per 1000 $\mu\text{mol s}^{-1} \text{m}^{-2}$
Output (in-water): 4.41 microamps per 1000 $\mu\text{mol s}^{-1} \text{m}^{-2}$

For use with LI-COR handheld meters and loggers:

Multiplier (in-air): -139.92 $\mu\text{mol s}^{-1} \text{m}^{-2}$ per microamp
Multiplier (in-water): -226.68 $\mu\text{mol s}^{-1} \text{m}^{-2}$ per microamp

For use with LI-COR 2291 (1210 ohm) Millivolt Adapter:

Multiplier (in-air): -115.64 $\mu\text{mol s}^{-1} \text{m}^{-2}$ per millivolt
Multiplier (in-water): -187.34 $\mu\text{mol s}^{-1} \text{m}^{-2}$ per millivolt

IMPORTANT: Read the appropriate instruction manual (www.licor.com/USM) before using this sensor.

IMPORTANT: It is recommended that sensors be recalibrated every two years after field deployment.

Calibration Technician: Aaron Deschane

Calibration standard used: Working standard lamp 811E.

Calibration traceable to the National Institute of Standards and Technologies (NIST) through NIST-calibration lamp number F-226.
(NBS Test No. 534/237689-86)



LI-COR Biosciences, Global Headquarters • 4647 Superior Street • Lincoln, NE 68504 USA
Phone: +1-402-467-3576 • Fax: +1-402-467-2819 • Toll-free: 800-447-3576 (USA & Canada)
envsales@licor.com • envsupport@licor.com • www.licor.com
Manual: www.licor.com/USM