



SEA-BIRD ELECTRONICS, INC.

13431 NE 20th St. Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

Pressure Test Certificate

Customer Pacific Marine Environmental Lab
Job Number 69513
Date 7/5/2012
Technician CJ

Serial Number 3928082-0991

Low Pressure (PSI) 45 PSI
Time (Minutes) 15 Minutes

High Pressure (PSI) 508 PSI
Time (Minutes) 30 Minutes

Pass ☒

Fail ☐

Comments

Replaced the main piston "O"-Rings.

