



# SEA-BIRD ELECTRONICS, INC.

13431 NE 20th St. Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 [www.seabird.com](http://www.seabird.com)

## Pressure Test Certificate

Customer Pacific Marine Environmental Lab

Job Number 62473

Date 1/24/2011

Technician DG

Serial Number 3928082-0990

Low Pressure (PSI) 50 PSI

Time (Minutes) 30 Minutes

High Pressure (PSI) \*N/A PSI

Time (Minutes) \*N/A Minutes

Pass ☒

Fail ☐

### Comments

Replaced the main piston "O"-Rings.

\*The unit is not designed for high pressure applications.

