

Certificate no:
Oxygen Optode 4330_4115_209060

Product: Oxygen Optode 4330
Serial no: 4115
Test Date: 20.12.2022

This is to certify that this product has been pressure tested:

Reference Equipment	Serial
Autoklav 800 bar	210005

Pressure Cycle Test:

Pressure (kPa)	Duration (Minutes)	Number of Cycles
60000	60	1

Pressure (kPa)	Duration (Minutes)
0	0

No irregularities were found during the test

Date:20.12.2022

Laila A. Skålnes

Laila Skålnes, Production Engineer