Pressure Test Certificate

Test Date: **2020-09-08**

Description: **SUNA-V2**

Sensor Information:

Model Number: **SUNA-V2**

Serial Number: **1467**

Pressure Test Protocol:

Low Pressure Test: **40**   PSI   Held For: **15**   Minutes

High Pressure Test: **150**   PSI   Held For: **15**   Minutes

Passed Test: **True**

Tested By: **dn**

High pressure is generally equal to the maximum depth rating of the instrument

Pressure

Time

Typical Test Profile