



**Sea-Bird Electronics, Inc.**

13431 NE 20<sup>th</sup> St. Bellevue, Washington 98005 USA  
www.seabird.com

Phone: (425) 643-9866

Fax: (425) 643-9954

Email: seabird@seabird.com

## Pressure Test Certificate

Test Date: **06/15/15**

Description: **SBE-43 DO Sensor**

### Sensor Information:

Model Number: **43**

Serial Number: **3200**

### Pressure Test Protocol:

Low Pressure Test: **40** PSI Held For: **15** Minutes

High Pressure Test: **10000** PSI Held For: **15** Minutes

Passed Test: **Yes**

Tested By: **RH**

**High pressure is generally equal to the maximum depth rating of the instrument**

