



Sea-Bird Scientific  
13431 NE 20<sup>th</sup> Street  
Bellevue, WA 98005  
USA

+1 425-643-9866  
seabird@seabird.com  
www.seabird.com

## Pressure Test Certificate

Test Date: **2022-09-15**

Description: **SBE-37 Microcat**

### Sensor Information:

Replaced the main piston "O"-Rings.

Model Number: **SBE-37**

Serial Number: **1842**

### Pressure Test Protocol:

Low Pressure Test: **40** PSI Held For: **15** Minutes

High Pressure Test: **500** PSI Held For: **30** Minutes

Passed Test: **True**

Tested By: **ND**

