Conductivity Calibration Report

Customer:	racine ivianne Environmental Lab				
Job Number:	62473	Dat	e of Report:	1/21/2	2011
Model Number	SBE 37SM	Ser	ial Number:	37SM280	82-2341
Conductivity sensors are normally calibrated 'as received', without cleaning or adjustments, allowing a determination of sensor drift. If the calibration identifies a problem or indicates cell cleaning is necessary, then a second calibration is performed after work is completed. The 'as received' calibration is not performed if the sensor is damaged or nonfunctional, or by customer request.					
An 'as received' calibration certificate is provided, listing the coefficients used to convert sensor frequency to conductivity. Users must choose whether the 'as received' calibration or the previous calibration better represents the sensor condition during deployment. In SEASOFT enter the chosen coefficients using the program SEACON. The coefficient 'slope' allows small corrections for drift between calibrations (consult the SEASOFT manual). Calibration coefficients obtained after a repair or cleaning apply only to subsequent data.					
'AS RECEIVED CALIBRATION' Performed Not Performed					Performed
Date: 1/21/2011		Drift since la	ast cal:	0.0000	PSU/month
Comments:					
'CALIBRATION AFTER CLEANING & REPLATINIZING' ☐ Performed ✓ Not Performed					
Date:		Drift since L	ast cal:		PSU/month
Comments:					
*Measured at 3.0	S/m				

Cell cleaning and electrode replatinizing tend to 'reset' the conductivity sensor to its original condition. Lack of drift in post-cleaning-calibration indicates geometric stability of the cell and electrical stability of the sensor circuit.