



# SEA-BIRD ELECTRONICS, INC.

13431 NE 20th St. Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 [www.seabird.com](http://www.seabird.com)

## Pressure Test Certificate

Customer Pacific Marine Environmental Lab

Job Number 83585

Date 3/6/2015

Technician CA

Serial Number 3944807-3769

Low Pressure (PSI) 45 PSI

Time (Minutes) 15 Minutes

High Pressure (PSI) 870 PSI

Time (Minutes) 30 Minutes

Pass ☒

Fail ☐

Comments

Replaced the main piston "O"-Rings.

