PO Box 518
620 Applegate St.
Philomath, OR 97370

(541) 929-5650
Fax (541) 929-5277
www.wetlabs.com

Chlorophyll WETStar Characterization

Date: January 17, 2017 S/N: WS3S-848P

Chlorophyll concentration expressed in µg/l can be derived using the equation:

CHL(µg/l) = Scale Factor x (Output - Clean Water Offset)

	Analog output	
Clean Water Offset (CWO)	0.068	V
Scale Factor (SF)	13.9	µg/l/V
Maximum Output	5.49	V
Resolution	0.15	mV
Ambient Characterization Temperature	22 ± 1	°C
Current Draw	30	mA @ 12V (typical)
12-hour Stability	0.08	mV/hr
Temperature Stability, 25–2 °C	0.22	mV/°C

Range	
15 µg/l	0
69 µg/l	X
150 µg/l	0

Definitions:

CWO: Clean Water Offset value obtained using pure filtered de-ionized water.

SF: Scale Factor is used to convert the fluorescence response of the instrument into chlorophyll-a concentration. Scale Factor is determined at WET Labs during a cross calibration using a liquid fluorescent standard and a reference fluorometer whose chlorophyll fluorescence response has been characterized in a laboratory using a mono-species lab culture of *Thalassiosira weissflogii* phytoplankton.

Maximum Output: Maximum signal output of the fluorometer.

Resolution: Standard deviation of 1 minute of clean water data, sampled once per second.

Ambient Characterization Temperature: Room temperature at time of characterization.

Current Draw: The amount of current the instrument uses for operation.

12-hour Stability: Deviation of output averaged over 12 hours.

Temperature Stability: Measured output variation per degree.

PO Box 518
620 Applegate St.
Philomath, OR 97370

(541) 929-5650
Fax (541) 929-5277
www.wetlabs.com

WETStar Calibration and Repairs

Date January 17, 2017 **Customer** NOAA

S/N# WS3S-848P **Repair Order** 32800

Standard Service

- Performed noise test: 1 sample/sec for 60 sec
- Performed stability test: 1 sample/min for 12 hrs
- Performed temperature test: 25–2 °C
- Performed saturation test
- Shake-tested unit
- Pressure-tested unit
- Updated unit's calibration sheet

Diagnosis

Evaluated Instrument and found no problems.

Repairs

Replaced the O-Rings.

Comments