



SEA-BIRD
SCIENTIFIC

SBE56 ECO Temperature Logger

Instrument Configuration

Instrument Serial Number: 56-13400
Instrument Firmware Version: 1.02
Communications Format: RS232

Installed Devices/Sensors

<i>Data Format</i>	<i>Measurement</i>	<i>Sensor Type</i>	<i>Serial Number</i>	<i>Rating</i>
Count	Temperature	Internal	N/A	N/A

Maximum Depth: **1500m**

CAUTION - The maximum deployment depth will be limited by the measurement range of the pressure sensor, if installed, an attached sensor, if installed, or the housing.



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 13400
CALIBRATION DATE: 20-Aug-23

SBE 56 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = -1.282155e-003
a1 = 3.551376e-004
a2 = -7.146041e-006
a3 = 2.118820e-007

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
-1.5000	20409668.3	-1.5000	0.0000
1.0000	18209251.8	0.9999	-0.0001
4.5000	15567114.5	4.5001	0.0001
8.0000	13353168.9	8.0000	0.0000
11.5000	11491493.3	11.4999	-0.0001
15.0000	9920695.8	15.0001	0.0001
18.5000	8591115.8	18.5000	-0.0000
22.0000	7462018.1	22.0000	-0.0000
25.5000	6500226.3	25.5000	0.0000
29.0000	5678476.4	29.0000	0.0000
32.5000	4974319.2	32.5000	-0.0000

n = Instrument Output (counts)

Temperature ITS-90 (°C) = $1/\{a0 + a1[\ln(n)] + a2[\ln^2(n)] + a3[\ln^3(n)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature

