

## a xylem brand

Sensing Foil Batch No:

1206

Certificate No:

3835 1033 1259

Product:

3835

Serial No:

1033

Calibration Date:

July 16, 2014

This is to certify that this product has been calibrated using the following instruments:

the product the over canonated abing the following moraliters.			
Fluke CHUB E-4	Serial No. A7C677		
Fluke 5615 PRT	Serial No. 849155		
Fluke 5615 PRT	Serial No. 802054		
Honeywell PPT	Serial No. 44074		
Calibration Bath model FNT 321-1-40	1		

Parameter: Internal Temperature:

Calibration points and readings:

	5			
Temperature (°C)	-		(	- "
Reading (mV)	-	-	-	-

Giving these coefficients

Index	0	1	2	3
TempCoef	2.11038E+01	-3.05522E-02	2.88430E-06	-4.16734E-09

<sup>\*</sup>Note: Temperature calibration NOT performed

Parameter: Oxygen:

	O2 Concentration	Air Saturation 0 - 120%	
Range:	0-500 μM <sup>1)</sup>		
Accuracy <sup>1)</sup> ;	< ±8μM or ±5%(whichever is greater)	±5%	
Resolution:	< 1 μΜ	< 0.4%	
Settling Time (63%):	< 25 seconds		

Calibration points and readings<sup>2)</sup>:

	Air Saturated Water	Zero Solution (Na <sub>2</sub> SO <sub>3</sub> )	
Phase reading (°) 3.06072E+01		6.39113E+01	
Temperature reading (°C) 1.01618E+01		1.98736E+01	
Air Pressure (hPa)	1.02100E+03		

Giving these coefficients

Index	0	I	2	3
PhaseCoef	-1.04092E+00	1.14415E+00	0.00000E+00	0.00000E+00

<sup>1)</sup> Valid for 0 to 2000m (6562ft) depth, salinity 33 - 37ppt

Date:

July 16, 2014

Signed: Jason B Baxter

Aanderaa Data Instruments

 $<sup>^{2)}</sup>$ The calibration is performed in fresh water and the salinity setting is set to: 0