



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

Pressure Test Certificate

Test Date: **2020-01-03**

Description: **SBE-16 SeaCat**

Sensor Information:

Model Number: **SBE-16**

Serial Number: **653**

Replaced 6-pin connector.
Replaced conductivity cell.
Replaced the main piston "O"-Rings.

Pressure Test Protocol:

Low Pressure Test: **40** PSI Held For: **15** Minutes

High Pressure Test: **870** PSI Held For: **30** Minutes

Passed Test: **True**

Tested By: **JK**

