Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

Pressure Test Certificate

Customer Pacific Marine Environmental Lab

Job Number 55436 <u>Date</u> 8/10/2009

DG Technician

| Serial Number | 4548614-0219 |
|---------------|--------------|
|---------------|--------------|

Low Pressure (PSI)

PSI 30

Time (Minutes)

30 Minutes

High Pressure (PSI)

*N/A **PSI**

Time (Minutes)

*N/A Minutes

Pass **V**

Fail

Comments

Replaced the main piston "O"-Rings.

*The unit is not designed for high pressure applications.

