13431 NE 20th St. Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

Pressure Test Certificate

<u>Customer</u> Pacific Marine Environmental Lab

Job Number 62186

<u>Date</u> 12/16/2010

Technician BLT

Serial Number 0	30701
-----------------	-------

Low Pressure (PSI) 50 PSI

Time (Minutes) 15 Minutes

High Pressure (PSI) 5000 PSI

<u>Time (Minutes)</u> 30 Minutes

Pass 🗸

Fail

Comments

Replaced the main piston "O"-Rings.

