Pressure Test Certificate

Test Date: **2020-01-08**

Description: **SBE-9plus CTD**

Sensor Information:

Model Number: **SBE-9P**

Serial Number: **91220-B**

Replaced the main piston "O"-Rings.
Replaced 6-pin connector.

Pressure Test Protocol:

Low Pressure Test: **40**   PSI   Held For: **15**   Minutes

High Pressure Test: **3000**   PSI   Held For: **30**   Minutes

Passed Test: **True**

Tested By: **nd**

High pressure is generally equal to the maximum depth rating of the instrument

Pressure

Time

Typical Test Profile