Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 2333
CALIBRATION DATE: 07-May-21

SBE 37 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

```
g = -9.907690e-001
h =  1.547243e-001
i = -2.149939e-004
j =  4.630302e-005

CPcor = -9.5700e-008
CTcor =  3.2500e-006
WBOTC =  9.2476e-006
```

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2532.27	0.00000	0.00000
1.0000	34.5953	2.95887	5050.86	2.95890	0.00003
4.5000	34.5754	3.26423	5241.53	3.26418	-0.00004
14.9999	34.5337	4.24059	5808.54	4.24060	0.00001
18.4999	34.5251	4.58388	5994.85	4.58391	0.00003
23.9999	34.5158	5.13883	6283.94	5.13880	-0.00003
29.0000	34.5107	5.65787	6542.45	5.65788	0.00001
32.5000	34.5069	6.02811	6720.57	6.02821	0.00010

f = Instrument Output(Hz) * sqrt(1.0 + WBOTC * t) / 1000.0

t = temperature (°C); p = pressure (decibars); δ = CTcor; ε = CPcor;

Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \varepsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity