

**Product:** Oxygen Optode 4330F

**Serial No:** 3157

**Date:** 07.02.2019

**Certificate No:** 1552741753157

This is to certify that this product has been pressure tested with the following instrument, and we confirm that no irregularities were found during the test:

Autoklav 800 bar – sn: 0210005

**Pressure readings:**

Pressure (Bar)	Pressure time (hour)
600	1

Date: 07 Feb 2019

Sign:

Laila A. Skålnes

Laila Skålnes, Production Engineer