

**Product:** Oxygen Optode 4831

**Serial No:** 682

**Date:** 06,03,2017

**Certificate No:** 128760260682

This is to certify that this product has been pressure tested with the following instrument, and we confirm that no irregularities were found during the test:


Autoklav 800 bar – sn: 0210005

**Pressure readings:**

Pressure (Bar)	Pressure time (hour)
600	1

Date: 23 Mar 2017

Sign:



Vidar Selsvik, Production Engineer