

Certificate no:
Oxygen Optode 4330F_4052_206229

Product: Oxygen Optode 4330F
Serial no: 4052
Test Date: 24.10.2022

This is to certify that this product has been pressure tested:

Reference Equipment	Serial
Autoklav 800 bar	210005

Pressure Cycle Test:

Pressure (kPa)	Duration (Minutes)	Number of Cycles
60000	60	1

No irregularities were found during the test

Date:24.10.2022

Laila A. Skålnes

Laila Skålnes, Production Engineer

Certificate no:
Oxygen Optode 4330F_4053_206230

Product: Oxygen Optode 4330F
Serial no: 4053
Test Date: 24.10.2022

This is to certify that this product has been pressure tested:

Reference Equipment	Serial
Autoklav 800 bar	210005

Pressure Cycle Test:

Pressure (kPa)	Duration (Minutes)	Number of Cycles
60000	60	1

No irregularities were found during the test

Date:24.10.2022

Laila A. Skålnes

Laila Skålnes, Production Engineer