

Job No.: R50937

Calibration Date: 05/30/23

Model Number: QSP2300

Serial Number: 70547

Operator: TPC

Standard Lamp: V-043(7/24/19)

Operating Voltage Range: 6 to 15 VDC (+)

Note: The QSP2300 output is a voltage that is proportional to the log of the incident irradiance. To calculate irradiance, use this formula:

$$\text{Irradiance} = \text{Calibration factor} * (10^{\wedge}\text{Light Signal Voltage} - 10^{\wedge}\text{Dark Voltage})$$

Dry Calibration Factor: 3.55E+12 quanta/cm<sup>2</sup>.sec per volt 5.90E-06  $\mu$ Einsteins/cm<sup>2</sup>.sec per volt  
Wet Calibration Factor: 6.27E+12 quanta/cm<sup>2</sup>.sec per volt 1.04E-05  $\mu$ Einsteins/cm<sup>2</sup>.sec per volt

#### Sensor Test Data and Results<sup>2)</sup>

Sensor Supply Current (Dark): 3.4 mA

Supply Voltage: 6 Volts

Lamp Integrated PAR Irradiance: 9.66E+15 quanta/cm<sup>2</sup>.sec 0.01605  $\mu$ Einsteins/cm<sup>2</sup>.sec  
Immersion Coefficient: 0.566

Nominal Filter OD	Expected Transmission	Calibrated Trans.	Sensor Voltage	Expected Voltage	Voltage % Error	Measured Trans.	Transmission Error (%)	Test Irrad. (quanta/cm <sup>2</sup> .sec)
No Filter	100%	100.00%	3.435	3.435	0%	100.00%	0.0	9.66E+15
0.3	50%	36.10%	2.998	2.993	0%	36.51%	-1.1	3.53E+15
0.5	32%	27.60%	2.886	2.876	0%	28.20%	-2.1	2.72E+15
1	10%	9.27%	2.418	2.402	1%	9.58%	-3.2	9.26E+14
2	1%	1.11%	1.514	1.480	2%	1.16%	-4.5	1.12E+14
3	0.10%	0.05%	0.358	0.163	55%	0.05%	14.7	4.55E+12
RG780	0.00%	0.00%	0.006	0.006	-2%	0.00%	-100.0	4.69E+10

Dark Before: 0.006 Volts  
Light - No Filter Hldr.: 3.435 Volts  
Dark After - NFH: 0.006 Volts  
Average Dark 0.0058 Volts

#### Notes:

1. Annual calibration is recommended.

2) This section is for internal use and for more advanced analysis.