Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 1853
CALIBRATION DATE: 04-May-21

SBE 37 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

```
g = -9.873843e-001
h =  1.360640e-001
i = -2.419828e-004
j =  4.215634e-005

CPcor = -9.5700e-008
CTcor =  3.2500e-006
WBOTC =  1.6066e-006
```

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2697.22	0.00000	0.00000
1.0000	34.6409	2.96240	5389.45	2.96244	0.00005
4.4999	34.6222	3.26820	5593.22	3.26815	-0.00005
15.0000	34.5815	4.24585	6199.04	4.24583	-0.00002
18.5000	34.5739	4.58967	6398.11	4.58967	-0.00000
24.0000	34.5664	5.14555	6707.09	5.14559	0.00005
29.0000	34.5638	5.66560	6983.28	5.66557	-0.00003
32.5001	34.5637	6.03691	7173.55	6.03649	-0.00043

f = Instrument Output(Hz) * sqrt(1.0 + WBOTC * t) / 1000.0

t = temperature (°C); p = pressure (decibars); δ = CTcor; ε = CPcor;

Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \varepsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity