Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 1851
CALIBRATION DATE: 24-Apr-19

SBE 37 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

```
g = -1.027775e+000
h =  1.404145e-001
i = -1.448131e-004
j =  3.462100e-005

CPcor = -9.5700e-008
CTcor =  3.2500e-006
WBOTC =  2.6784e-006
```

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2706.73	0.00000	0.00000
1.0000	34.7911	2.97402	5334.50	2.97403	0.00002
4.5000	34.7711	3.28088	5534.31	3.28086	-0.00002
15.0000	34.7282	4.26195	6128.75	4.26193	-0.00002
18.5000	34.7188	4.60683	6324.15	4.60684	0.00001
23.9999	34.7079	5.16427	6627.45	5.16430	0.00003
29.0000	34.7005	5.68548	6898.58	5.68547	-0.00002
32.5000	34.6941	6.05709	7085.29	6.05694	-0.00015

f = Instrument Output(Hz) * sqrt(1.0 + WBOTC * t) / 1000.0

t = temperature (°C); p = pressure (decibars); δ = CTcor; ε = CPcor;

Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \varepsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity