Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 1804
CALIBRATION DATE: 28-May-21

SBE 37 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

```
g = -9.757384e-001
h =  1.382889e-001
i = -2.055019e-004
j =  3.876989e-005

CPcor = -9.5700e-008
CTcor =  3.2500e-006
WBOTC =  2.7581e-006
```

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2658.81	0.00000	0.00000
0.9999	34.6384	2.96219	5336.17	2.96220	0.00001
4.5000	34.6183	3.26788	5538.52	3.26787	-0.00001
14.9999	34.5765	4.24529	6140.04	4.24531	0.00002
18.4999	34.5682	4.58898	6337.64	4.58897	-0.00001
24.0000	34.5593	5.14461	6644.33	5.14460	-0.00000
29.0001	34.5544	5.66424	6918.46	5.66424	-0.00000
32.5000	34.5513	6.03498	7107.35	6.03499	0.00000

f = Instrument Output(Hz) * sqrt(1.0 + WBOTC * t) / 1000.0

t = temperature (°C); p = pressure (decibars); δ = CTcor; ε = CPcor;

Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \varepsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity