

Product: Oxygen Optode 4831

Certificate No: 154674260800

Serial No: 800

Date: 08.01.2019

This is to certify that this product has been pressure tested with the following instrument, and we confirm that no irregularities were found during the test:


Autoklav 800 bar – sn: 0210005

Pressure readings:

Pressure (Bar)	Pressure time (hour)
600	1

Date: 16 Jan 2019

Sign:



Vidar Selsvik, Production Engineer

Product: Oxygen Optode 4831

Certificate No: 154675260801

Serial No: 801

Date: 08.01.2019

This is to certify that this product has been pressure tested with the following instrument, and we confirm that no irregularities were found during the test:


Autoklav 800 bar – sn: 0210005

Pressure readings:

Pressure (Bar)	Pressure time (hour)
600	1

Date: 16 Jan 2019

Sign:



Vidar Selsvik, Production Engineer