



## Chlorophyll WETStar Characterization

Date: January 8, 2014

S/N: WS3S-848P

Chlorophyll concentration expressed in  $\mu\text{g/l}$  can be derived using the equation:

$$\text{CHL}(\mu\text{g/l}) = \text{Scale Factor} \times (\text{Output} - \text{Clean Water Offset})$$

Clean Water Offset (CWO)	Analog output 0.067 V
Scale Factor (SF)	15.7 $\mu\text{g/l/V}$
Maximum Output	5.49 V
Resolution	0.18 mV
Ambient Characterization Temperature	22 $\pm$ 1°C
Current Draw	30 mA @ 12V (typical)
12-hour Stability	0.13 mV/hr
Temperature Stability, 25–2 °C	0.10 mV/°C

Range	
15 $\mu\text{g/l}$	0
77 $\mu\text{g/l}$	X
150 $\mu\text{g/l}$	0

### Definitions:

**CWO:** Clean Water Offset value obtained using pure filtered de-ionized water.

**SF:** Scale Factor is used to convert the fluorescence response of the instrument into chlorophyll-a concentration. Scale Factor is determined at WET Labs during a cross calibration using a liquid fluorescent standard and a reference fluorometer whose chlorophyll fluorescence response has been characterized in a laboratory using a mono-species lab culture of *Thalassiosira weissflogii* phytoplankton.

**Maximum Output:** Maximum signal output of the fluorometer.

**Resolution:** Standard deviation of 1 minute of clean water data, sampled once per second.

**Ambient Characterization Temperature:** Room temperature at time of characterization.

**Current Draw:** The amount of current the instrument uses for operation.

**12-hour Stability:** Deviation of output averaged over 12 hours.

**Temperature Stability:** Measured output variation per degree.

PO Box 518  
620 Applegate St.  
Philomath, OR 97370



(541) 929-5650  
Fax (541) 929-5277  
[www.wetlabs.com](http://www.wetlabs.com)

---

## WETStar Calibration and Repairs

**Date** January 8, 2014      **Customer** NOAA

**S/N#** WS3S-848P      **Repair Order** 21594

---

### Standard Service

- Performed noise test: 1 sample/sec for 60 sec
- Performed stability test: 1 sample/min for 12 hrs
- Performed temperature test: 25–2 °C
- Performed saturation test
- Shake-tested unit
- Pressure-tested unit
- Updated unit's calibration sheet

### Diagnosis

Socket one on Bulkhead Connector is loose.  
Red Filter was delaminated.

### Repairs

Replaced the Bulkhead Connector, Red Filter and O-Rings.

### Comments

WETStar was re-calibrated with 25ppb Uranine.  
Installed new 6 Pin Bulkhead Connector.  
Analog Patch Cable Part #: EXA-210317