

# CALIBRATION CERTIFICATE

a xylem brand

**Layout No:**  
**Range:** 0 - 20 MPa

**Product:** Pressure Sensor 3815D  
**Serial No:** 401  
**Calibration Date:** November 20, 2014

## Calibration

The calibration of each sensor is carried out at the factory and the calibration is valid for all following data handling. The form below is filled in only where necessary, depending upon which type of sensor it concerns.

Pressure		Coefficients (kPa)	
Bar	Reading N		
Air	34	A	-5.858E-01
50	279	B	2.026E-02
100	522	C	3.241E-07
150	766	D	0.000E+00
200	1004		

Date:  
November 20, 2014

Sign: Jason B Baxter