



AANDERAA
INSTRUMENTS

5852 Bergen, Norway. Tel. + 47 55 10 99 00

212014

Test and Specification Sheet

Recording Current Meter,

RCM 9, Mk II ☐

RCM 11 ☒

Serial No. 62

Main Components

Component	Serial No.	Remarks
Electronic Board 3623	807	Reference reading: 735
Data Storage Unit 2990	12439	Standard <input checked="" type="checkbox"/> Extended <input type="checkbox"/>
Doppler Current Sensor 3620 /3820	62	
Temperature Sensor 3621	823	
Conductivity Cell 3640 /3819	28	
Pressure sensor 3815 E	156	
Turbidity sensor 3612		
Oxygen Sensor 3675		

Mechanical Checks:

Sensors fixed in correct position ☒
Wire harness, screws and sensor plugs ☒
Epoxy coating intact ☒
Zinc anode installed ☒
Clean and inspect O-ring groove ☒

Performance Tests of complete instrument:

Tune the transducer so that the acoustic output of 16.384 KHz is equal to the output of 15.7 KHz ☐
Current consumption at continuous operation, maximum 100 mA 99.6 mA
Current consumption between measurements at 120 min. interval. Maximum 1.0 mA average 0.65 mA
Field test 8 hours, 5 min. interval recording in DSU 2990 ☒
Check operation with Test Unit 3731, -5°C to +35°C, (all channels tested, 16 hour run, data stored in DSU 2990) ☒
Check remote start, PDC-4 output and external powering ☒
Electrical isolation between system ground and Top end-plate ☒
Test of Operation in Burst Mode, 2 minutes interval ☒

Date 19/4-01 Sign J. Heltn

Final Check prior to Shipment:

Doppler Current Sensor is tested with Test Unit 3731 ☒
Temperature readings correspond to room temperature ☒
Conductivity Sensor reads correct with sea water loop ☒
Pressure Sensor gives correct reading at air pressure ☒
Turbidity reading increases when a reflector is placed 20cm in front of it ☐
The oxygen sensor reads maximum in air ☐

Optional Sensor checked by: (sign)

Set temperature range switch to "Low" and conductivity range to 0-74 mS ☒
Set interval switch to 10 minutes and turn channel selector switch to OFF position ☒
Erased DSU installed ☒
Install fresh battery, type Alkaline 3614, Open loop voltage: >9.5 Voltage with 100ohms load: >9.2 ☒
Inspect O-ring groove and clean and grease O-ring ☒
Check that the pressure sensor is oil filled ☒

Date 19/4-01 Sign [Signature]



**AANDERAA
INSTRUMENTS**

5852 Bergen, Norway. Tel. + 47 55 10 99 00

Calibration Sheet

Recording Current Meter,

RCM 9, Mk II ☐

RCM 11 ☒

Serial No.: **62**

Reference(channel 1): **735**

The calibration coefficients listed below are valid for the sensors with the following serial numbers

Sensor	Range	Serial No.
Doppler Current Sensor 3020 /3820		62
Temperature Sensor 3621		823
Conductivity cell 3619 /3819		28
Pressure sensor 3815 E	0-60 MPa	156
Turbidity sensor 3612		
Oxygen Sensor 3675		

Calibration Coefficients:

Ch.No.	Parameter	A	B	C	D	Unit
1	Reference	0	1.000E+00	0	0	
2	Current Speed	0	2.933E-01	0	0	cm/s
3	Current Direction	0	3.516E-01	0	0	Deg.M
4	Temperature					
	Wide range	-6,036 E-01	3,416 E-02	-6.292E-06	4.800E-09	Deg.C
	Low range	-2,682 E+00	2,407 E-02	-2.238E-06	2.056E-09	Deg.C
	High range	9,815 E+00	2,390 E-02	-5.941E-07	2.820E-09	Deg.C
	Arctic range	-2,956 E+00	8,948 E-03	-3.476E-07	1.134E-10	Deg.C
5	Conductivity					
	0-74mS/cm	6,745 E-02	6,745 E-02	0	0	mS/cm
	24-38mS/cm	2,224 E+01	1,521 E-02	0	0	mS/cm
	0-2mS/cm			0	0	mS/cm
6	Pressure	-1,771 E+00	6,272 E-02	-8,975 E-07	0	kPa/MPa
7	Turbidity					NTU
	Oxygen			0	0	% *

*

Oxygen saturation with respect to nominal air pressure (1013.2 hPa)

To obtain the density of oxygen in mg/l, see calibration sheet for the Oxygen Sensor, Form 533

Date: **19/4-2001** Sign: **Pau Sagmes**