



# SEA-BIRD ELECTRONICS, INC.

13431 NE 20th St. Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

## Pressure Test Certificate

Customer Pacific Marine Environmental Lab  
Job Number 62186  
Date 2/1/2011  
Technician DG

Serial Number 37SM36048-3764

Low Pressure (PSI) 50 PSI  
Time (Minutes) 15 Minutes

High Pressure (PSI) 1500 PSI  
Time (Minutes) 30 Minutes

Pass ☒

Fail ☐

Comments

Replaced the main piston "O"-Rings.

