



Sea-Bird Electronics, Inc.
13431 NE 20th Street
Bellevue, WA 98005 United States

Phone

+1-425-643-9866

www.seabird.com

1005518695

29-JUN-2023

320593155

SERVICE REPORT

Service Request

Date

Sales Order

PO Number:

CUSTOMER INFORMATION

Name: NOAA - PMEL PACIFIC MARINE ENVIRO LAB

Account : 40279844

DAVID STRAUZ

DAVID.STRAUSZ@NOAA.GOV

206-526-4510

Bill To Address

NOAA - PMEL PACIFIC MARINE ENVIRO LAB
7600 SAND POINT WAY NE

SEATTLE,WA,98115-6349,US

Ship To Address

NOAA - PMEL PACIFIC MARINE ENVIRO LAB
SUPPLY CENTER / RECEIVING
7600 SAND POINT WAY NE, BLDG 8
SEATTLE,WA,98115,US

PRODUCT INFORMATION

Item: 43.2111

Item Description: SBE43, 1/2 MIL, 7000m, XSG, STD 43 PLENUM

Serial: 43-3200

Special Notes

Services Requested:

CAL_43

Services Performed:

Perform initial diagnostic evaluation.

Performed "POST" cruise calibration.

Item	Item Description	Qty
CAL_43	SBE 43 dissolved oxygen sensor calibration	1

Unbilled Items

Item	Item Description	Qty



Sea-Bird Electronics, Inc.
13431 NE 20th Street
Bellevue, WA 98005 United States

Phone

+1-425-643-9866

www.seabird.com

1005518695

29-JUN-2023

320593155

SERVICE REPORT

Service Request

Date

Sales Order

PRODUCT INFORMATION

Item: 43.LEGACY

Item Description: (LEGACY) SBE 43 Oxygen Sensor

Serial: 431875

Special Notes

Services Requested:

CAL_43

Services Performed:

Perform initial diagnostic evaluation.

Performed "POST" cruise calibration.

Item	Item Description	Qty
CAL_43	SBE 43 dissolved oxygen sensor calibration	1

Unbilled Items

Item	Item Description	Qty



Sea-Bird Electronics, Inc.
13431 NE 20th Street
Bellevue, WA 98005 United States

Phone

+1-425-643-9866

www.seabird.com

1005518695

29-JUN-2023

320593155

SERVICE REPORT

Service Request

Date

Sales Order

PRODUCT INFORMATION

Item: 03P.LEGACY

Item Description: (LEGACY) SBE 03Plus Temperature Sensor

Serial: 03P2786

Special Notes

Services Requested:

CAL_3

Services Performed:

Perform initial diagnostic evaluation.

Performed "POST" cruise calibration.

Item	Item Description	Qty
CAL_3	SBE 3 Temperature sensor calibration; 11 points -1 to 31deg.C	1

Unbilled Items

Item	Item Description	Qty



Sea-Bird Electronics, Inc.
13431 NE 20th Street
Bellevue, WA 98005 United States

Phone

+1-425-643-9866

www.seabird.com

1005518695

29-JUN-2023

320593155

SERVICE REPORT

Service Request

Date

Sales Order

PRODUCT INFORMATION

Item: 04C.LEGACY

Item Description: (LEGACY) SBE 04C Conductivity Sensor

Serial: 042490

Special Notes

Services Requested:

CAL_4C

Services Performed:

Perform initial diagnostic evaluation.

Performed "POST" cruise calibration.

Item	Item Description	Qty
CAL_4C	SBE 4C Conductivity Calibration for CTD sensors (includes 6-point 35 PSU post deployment cal., and C&P with 6-point 35 PSU re-cal, if necessary)	1

Unbilled Items

Item	Item Description	Qty
23548	Zinc anode ring (sensors with tubing connector)	1



Sea-Bird Electronics, Inc.
13431 NE 20th Street
Bellevue, WA 98005 United States

Phone

+1-425-643-9866

www.seabird.com

1005518695

29-JUN-2023

320593155

SERVICE REPORT

Service Request

Date

Sales Order

PRODUCT INFORMATION

Item: 04C.LEGACY

Item Description: (LEGACY) SBE 04C Conductivity Sensor

Serial: 043994

Special Notes

Services Requested:

CAL_4C

Services Performed:

Perform initial diagnostic evaluation.

Performed "POST" cruise calibration.

Item	Item Description	Qty
CAL_4C	SBE 4C Conductivity Calibration for CTD sensors (includes 6-point 35 PSU post deployment cal., and C&P with 6-point 35 PSU re-cal, if necessary)	1

Unbilled Items

Item	Item Description	Qty



Sea-Bird Electronics, Inc.
13431 NE 20th Street
Bellevue, WA 98005 United States

Phone

+1-425-643-9866

www.seabird.com

1005518695

29-JUN-2023

320593155

SERVICE REPORT

Service Request

Date

Sales Order

PRODUCT INFORMATION

Item: 04C.LEGACY

Item Description: (LEGACY) SBE 04C Conductivity Sensor

Serial: 040334

Special Notes

Services Requested:

CAL_4C

Services Performed:

Perform initial diagnostic evaluation.

Performed "POST" cruise calibration.

Item	Item Description	Qty
CAL_4C	SBE 4C Conductivity Calibration for CTD sensors (includes 6-point 35 PSU post deployment cal., and C&P with 6-point 35 PSU re-cal, if necessary)	1

Unbilled Items

Item	Item Description	Qty