PO Box 518 620 Applegate St. Philomath, OR 97370



(541) 929-5650 Fax (541) 929-5277 www.wetlabs.com

## **Chlorophyll WETStar Characterization**

Date: February 29, 2012 S/N: WS3S-742P

Chlorophyll concentration expressed in µg/l can be derived using the equation:

 $CHL(\mu g/I) = Scale Factor x (Output - Clean Water Offset)$ 

Analog output
Clean Water Offset (CWO)

Scale Factor (SF)

Analog output

0.065 V @

5.5 μg/l/V @

Maximum Output 5.18 V @ Resolution 0.36 mV Ambient Characterization Temperature  $22 \pm 1^{\circ}$ C

Current Draw 40 mA @ 12V (typical)

12-hour Stability 0.28 mV/hr Temperature Stability, 25–2  $^{\circ}$  0.17 mV/ $^{\circ}$ 

Range	
15 μg/l	0
28 μg/l	Χ
150 µg/l	0

#### Definitions:

CWO: Clean Water Offset value obtained using pure filtered de-ionized water.

**SF**: Scale Factor is used to convert the fluorescence response of the instrument into chlorophyll-a concentration. Scale Factor is determined at WET Labs during a cross calibration using a liquid fluorescent standard and a reference fluorometer whose chlorophyll fluorescence response has been characterized in a laboratory using a mono-species lab culture of *Thalassiosira weissflogii* phytoplankton.

10/3/07

Maximum Output: Maximum signal output of the fluorometer.

**Resolution:** Standard deviation of 1 minute of clean water data, sampled once per second. **Ambient Characterization Temperature:** Room temperature at time of characterization.

**Current Draw:** The amount of current the instrument uses for operation.

12-hour Stability: Deviation of output averaged over 12 hours.

Temperature Stability: Measured output variation per degree.

WS3S-742P wkbk.xls Revision I

PO Box 518 620 Applegate St. Philomath, OR 97370



(541) 929-5650 Fax (541) 929-5277 www.wetlabs.com

# **WETStar Calibration and Repairs**

Date February 29, 2012 Customer NOAA

S/N# WS3S-742P Repair Order 13997

### **Standard Service**

- Performed noise test: 1 sample/sec for 60 sec
- Performed stability test: 1 sample/min for 12 hrs
- Performed temperature test: 25–2 ℃
- Performed saturation test
- Shake-tested unit
- Pressure-tested unit
- Updated unit's calibration sheet

### **Additional Repairs**

Parts Replaced: Bulkhead Connector and O-Rings.

#### **Comments**

WETStar was re-calibrated with 100ppb Uranine. Installed new 6 Pin Bulkhead Connector. Analog Patch Cable Part #: EXA-210317

WS3S-742P wkbk.xls Revision I 4/17/08