Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 3762
CALIBRATION DATE: 06-Apr-23

SBE 37 PRESSURE CALIBRATION DATA
1450 psia S/N 5751

COEFFICIENTS:

| | | | |
|---|---|---|---|
| PA0 = | 2.430573e-001 | PTCA0 = | -1.912977e+002 |
| PA1 = | 6.734028e-002 | PTCA1 = | 5.779511e-002 |
| PA2 = | -3.532263e-009 | PTCA2 = | 4.720649e-003 |
| | | PTCB0 = | 2.540563e+001 |
| | | PTCB1 = | 1.250000e-004 |
| | | PTCB2 = | 0.000000e+000 |

PRESSURE SPAN CALIBRATION

| PRESSURE (PSIA) | INSTRUMENT OUTPUT (counts) | TEMPERATURE (°C) | COMPUTED PRESSURE (PSIA) | RESIDUAL (%FSR) |
|---|---|---|---|---|
| 14.43 | 22.1 | 22.2 | 14.37 | -0.00 |
| 301.66 | 4289.5 | 22.2 | 301.63 | -0.00 |
| 588.61 | 8553.7 | 22.3 | 588.55 | -0.00 |
| 875.67 | 12822.5 | 22.3 | 875.66 | -0.00 |
| 1162.92 | 17096.2 | 22.3 | 1162.96 | 0.00 |
| 1450.03 | 21367.8 | 22.3 | 1449.99 | -0.00 |
| 1162.93 | 17096.0 | 22.3 | 1162.95 | 0.00 |
| 875.90 | 12826.6 | 22.4 | 875.93 | 0.00 |
| 588.73 | 8556.9 | 22.4 | 588.77 | 0.00 |
| 301.65 | 4289.5 | 22.4 | 301.63 | -0.00 |
| 14.42 | 24.3 | 22.7 | 14.51 | 0.01 |

THERMAL CORRECTION

| TEMP (°C) | INSTRUMENT OUTPUT (counts) |
|---|---|
| 32.50 | 40.93 |
| 29.00 | 39.86 |
| 24.00 | 38.31 |
| 18.50 | 36.65 |
| 15.00 | 36.06 |
| 4.50 | 34.64 |
| 1.00 | 34.09 |

| TEMPERATURE (°C) | SPAN |
|---|---|
| -5.00 | 25.41 |
| 35.00 | 25.41 |

x = instrument output - PTCA0 - PTCA1 * t - PTCA2 * $t^2$

n = x * PTCB0 / (PTCB0 + PTCB1 * t + PTCB2 * $t^2$)

pressure (PSIA) = PA0 + PA1 * n + PA2 * $n^2$

Residual (%FSR) = (computed pressure - true pressure) * 100 / Full Scale Range