

AANDERAA *Summary of Calibration Certificates*

Page 1 of 1

Recording Current Meter: **RCM 9 MkII** Serial No: **1199**

Electronic Board: **3970** Serial No: **298**

Reference reading: **510**

Calibration Dates: See the individual Calibration Certificates

The calibration coefficients listed below are valid for sensors with the following serial numbers:

Sensor	Type	Serial No.	Range
Doppler Current Sensor	3920	920	
Temperature Sensor	3621	1845	
Pressure Sensor	3815 B	616	0 – 3500 kPa

Calibration Coefficients:

Ch. No.	Parameter	A	B	C	D	Unit
1	Reference	0	1.000E+00	0	0	–
2	Current Speed	0	2.933E-01	0	0	cm/s
3	Current Direction	0	3.516E-01	0	0	Deg. M
4	Temperature Range					
	Arctic	-3.147E+00	8.980E-03	-3.476E-07	1.134E-10	Deg. C
	High	9.686E+00	2.399E-02	-5.941E-07	2.820E-09	Deg. C
	Low	-2.856E+00	2.411E-02	-2.238E-06	2.056E-09	Deg. C
	Wide	-7.616E-01	3.424E-02	-6.292E-06	4.800E-09	Deg. C
	Other					Deg. C
5	Conductivity Range					
6	Pressure	-1.578E+01	3.438E+00	5.822E-05	0	kPa
7						
8						

Date: 01 October 2004

Sign. **Roar Sognnes**
Calibration Engineer

PO BOX 160 NESTTUN
5852 BERGEN, NORWAY
Fax. +47 55 10 99 10

NESTTUNBREKKEN 97
5221 NESTTUN, NORWAY
Tel. +47 55 10 99 00

E-mail: info@aanderaa.no
Visit our Web Site at:
<http://www.aanderaa.com>

Enterprise No: NO 943 521638
Bank Account No: 5210.05.31267
SWIFT code: DNBANOKK