

PO Box 518
620 Applegate St.
Philomath OR 97370



SEA-BIRD
SCIENTIFIC

(541) 929-5650
Fax (541) 929-5277
www.seabird.com

WETStar Calibration and Repairs

S/N# WS3S-805P

Standard Service

Performed noise test: 1 sample/sec for 60 sec.
Performed stability test: 1 sample/min for 12 hrs.
Performed temperature test: 25-2 °C.
Performed saturation test.
Shake-tested unit.
Pressure-tested unit.
Updated unit's calibration sheet.

Diagnosis

Perform Standard Services.

Repairs

Standard service performed.
Replaced O-rings.

Comments

None.