Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

Pressure Test Certificate

Test Date: **2024-04-26**

Description: **SBE-39 Temperature/Pressure Sensor**

Sensor Information:

Model Number: **SBE-39**

Serial Number: **1420**

Pressure Test Protocol:

Low Pressure Test: **40**   PSI   Held For: **15**   Minutes

High Pressure Test: **500**   PSI   Held For: **15**   Minutes

Passed Test: **True**

Tested By: **JL**

High pressure is generally equal to the maximum depth rating of the instrument

Pressure

Time

Typical Test Profile