



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 3767
CALIBRATION DATE: 12-Apr-23

SBE 37 PRESSURE CALIBRATION DATA
1450 psia S/N 5756

COEFFICIENTS:

PA0 = 6.779824e-001
PA1 = 6.918869e-002
PA2 = -3.275093e-009

PTCA0 = -1.707943e+002
PTCA1 = 4.656467e-001
PTCA2 = 1.449014e-004
PTCB0 = 2.469425e+001
PTCB1 = -5.500000e-004
PTCB2 = 0.000000e+000

PRESSURE SPAN CALIBRATION

THERMAL CORRECTION

PRESSURE (PSIA)	INSTRUMENT OUTPUT (counts)	TEMPERATURE (°C)	COMPUTED PRESSURE (PSIA)	RESIDUAL (%FSR)	TEMP (°C)	INSTRUMENT OUTPUT (counts)
14.52	38.6	20.9	14.49	-0.00	32.50	54.63
314.81	4358.5	21.2	313.45	-0.09	29.00	53.25
614.81	8712.4	21.2	614.64	-0.01	24.00	50.99
914.80	13051.5	21.3	914.68	-0.01	18.50	48.15
1214.83	17392.0	21.3	1214.70	-0.01	15.00	46.32
1464.77	21011.1	21.3	1464.76	-0.00	4.50	41.61
1214.66	17393.7	21.3	1214.82	0.01	1.00	40.02
914.67	13052.9	21.2	914.78	0.01	TEMPERATURE (°C)	SPAN
614.65	8713.7	21.3	614.73	0.01		
314.69	4378.2	21.3	314.81	0.01		
14.53	39.0	21.3	14.51	-0.00		
					-5.00	24.70
					35.00	24.68

$$x = \text{instrument output} - \text{PTCA0} - \text{PTCA1} * t - \text{PTCA2} * t^2$$

$$n = x * \text{PTCB0} / (\text{PTCB0} + \text{PTCB1} * t + \text{PTCB2} * t^2)$$

$$\text{pressure (PSIA)} = \text{PA0} + \text{PA1} * n + \text{PA2} * n^2$$

$$\text{Residual (\%FSR)} = (\text{computed pressure} - \text{true pressure}) * 100 / \text{Full Scale Range}$$

