



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

Pressure Test Certificate

Test Date: **2024-08-15**

Description: **SUNA-V2**

Sensor Information:

Replaced the main piston "O"-Rings.

Model Number: **SUNA-V2**

Serial Number: **0598**

Pressure Test Protocol:

Low Pressure Test: **40** PSI Held For: **15** Minutes

High Pressure Test: **150** PSI Held For: **15** Minutes

Passed Test: **True**

Tested By: **AP**

High pressure is
generally equal
to the maximum
depth rating of
the instrument

