

Reference excerpt 01-032

Source: arXiv:1604.05171v1 | License: CC BY 4.0 | Word window: 1101-1360

JVP trace of pressure in the IJV, measured in G, together with the ECG trace. The interval t_{aQP} is represented over the curves a and QP. In figure b) it is shown the pressure along the z-axes at instant t_{QP} . pressure in the IJV, measured in G, can be obtained simultaneously with the acquisition of the ECG trace [Sisini et al.(2016)]. On this trace, that reports both JVP and ECG waves, the time interval t_{aQP} between t_P Q and t_a is given by: $t_{aQP} = t_a - t_{QP} = IG/c$ (4) where t_a is the instant corresponding the a wave. The parameter t_{aQP} is the time interval between the instant when the pressure is maximum at point G (instant t_a) and the instant when the pressure is maximum at point RA (instant t_{QP}). METHODS The applicability of the relationship expressed in Eq. 4, can be tested by comparing the instantaneous velocity ($w(t)$) of the blood in the IJV calculated according to the equation shown above, with the instantaneous blood velocity ($w_D(t)$) measured with a Doppler scanner. However, the data and the results presented here are for illustration only and are not to be considered the result of a scientific study. Pressure waves velocity calculation The delay t_{aQP} is measured over the JVP+ECG trace as shown in Fig. 2. The distance IG is measured on the volunteers chest. The velocity c is the ratio between IG and t_{aQP} . Compliance calculation From the Moens-Korteweg equation we obtain the compliance per