

Reference excerpt 02-015

Source: arXiv:1604.05177v1 | License: CC BY 4.0 | Word window: 532-791

this work, the video clip of the IJV was analyzed frame by frame to find the sonograms in which the cursor of the ECG was at one of the events P, R or T (see Fig. 1). The time instants relating to each event were denoted t_{Pi} , t_{Ri} and t_{Ti} , where i indicates the i -th cardiac cycle between those traces. JVP quantitative evaluation The waves a, c, x, v and y and the intervals a and v are the parameters that characterize the JVP (see Fig. 2). In this paper, these parameters are denoted in terms of the values assumed by the CSA during the cardiac cycle. The values of the CSA of the IJV in correspondence of the waves a, c, x, v and y, during the i -th cycle are indicated by the parameters a_i , c_i , x_i , v_i and y_i . These parameters are detected on the JVP trace by means of an algorithm based on the following definitions: $a_1 = \max(CSA(t)) : 0 < t < T$ $c_i = \max(CSA(t)) : t_{yi-1} + t_{ya} - T_c < t < t_{yi-1} + t_{ya} + T_c$ $c_i = \max(CSA(t)) : t_{ai} + t_{ac} - T_c < t < t_{ai} + t_{ac} + T_c$ $x_i = \min(CSA(t)) : t_{ai} + t_{ax} - T_c < t < t_{ai} + t_{ax} + T_c$ $v_i = \max(CSA(t)) : t_{xi} + t_{xv} - T_c < t < t_{xi} + t_{cx} + T_c$ $y_i = \min(CSA(t)) : t_{vi} + t_{vy} - T_c$