

Reference excerpt 02-033

Source: arXiv:1604.05177v1 | License: CC BY 4.0 | Word window: 1215-1474

standard deviation of main JVP parameters are presented. a x v a v t ax tvy HR (cm2) (cm2) (cm2) (cm2) (s) (s) Bpm 1.57 0.18 1.29 0.16 1.33 0.18 0.29 0.04 0.04 0.02 0.33 0.06 0.11 0.05 61 Table 2: Example of mean value and standard deviation of of time intervals t aR and t xT averaged over several cardiac cycles. taR txT taP (s) (s) (s) 0,03 0,02 0,12 0,06 0.14 0.05 Figure 2: The figure shows an illustrative JVP trace during a cardiac cycle together with the ECG trace. The waves a, c, x, v and y of the JVP and the events P, R and T of the ECG are indicated. The average and the standard deviation of these three parameters is calculated. The mean values and standard deviations of the parameters taR, txT and taP are presented in Table 2. Figure 3: Examples of CSA and ECG trace are shown. The waves a, x, v and y are correctly detected by an algorithm. The P, R and T events of the ECG trace are indicated. Figure 4: Examples of CSA and ECG trace are shown. The x and v and y are not correctly detected. Correct waves are indicated. The P, R and T events of the ECG are indicated. 3 DISCUSSION This note describes a methodology to extrapolate some interesting quantitative parameters from the IJV CSA trace, which is an analogue of the JVP. There are two important aspects that must be considered: i)