

Reference excerpt 02-034

Source: arXiv:1604.05177v1 | License: CC BY 4.0 | Word window: 1253-1512

0.11 0.05 61 T able 2: Example of mean value and standard deviation of of time inter vals t aR and t xT averaged over several cardiac cycles. taR txT taP (s) (s) (s) 0,03 0,02 0,12 0,06
0.14 0.05 Figure 2: The gure shows an illustrative JVP trace during a cardiac cycle together with the ECG trace. The ve waves a, c, x, v and y of the JVP and the events P, R and T of the ECG are indicated. The average and the standard deviation of these three parame - ters is calculated. The mean values and standard deviations of the parameters taR, txT and taP are presented in Table 2. Figure 3: Examples of CSA and ECG trace are shown. The waves a, x, v e y are correctly detected by an algorithm. The P, R e T events of the ECG trace are indicated. Figure 4: Examples of CSA and ECG trace are shown. The x and v e y are not correctly detected. Correct waves are indicated. The P, R e T events of the ECG are indicated. 3 DISCUSSION This note describes a methodology to extrapolate some inter - esting quantitative parameters from the IJV CSA trace, whic h is an analogue of the JVP. There are two important aspects that must be considered: i) the contouring of the IJV on each sonogram of the video-clip must be as accurate as possible. This means that the algorithm presented in [Sisini et al.(20 15)] must be improved to achieve an accuracy close to 100