

Reference excerpt 02-035

Source: arXiv:1604.05177v1 | License: CC BY 4.0 | Word window: 1291-1550

0.14 0.05 Figure 2: The figure shows an illustrative JVP trace during a cardiac cycle together with the ECG trace. The waves a, c, x, v and y of the JVP and the events P, R and T of the ECG are indicated. The average and the standard deviation of these three parameters is calculated. The mean values and standard deviations of the parameters ta_R , tx_T and ta_P are presented in Table 2. Figure 3: Examples of CSA and ECG trace are shown. The waves a, x, v and y are correctly detected by an algorithm. The P, R and T events of the ECG trace are indicated. Figure 4: Examples of CSA and ECG trace are shown. The x and v and y are not correctly detected. Correct waves are indicated. The P, R and T events of the ECG are indicated. 3 DISCUSSION This note describes a methodology to extrapolate some interesting quantitative parameters from the IJV CSA trace, which is an analogue of the JVP. There are two important aspects that must be considered: i) the contouring of the IJV on each sonogram of the video-clip must be as accurate as possible. This means that the algorithm presented in [Sisini et al.(2015)] must be improved to achieve an accuracy close to 100 %, while the current one is about 95 %. ii) The wave detection algorithm must be improved. In fact it is not always sufficient to identify the waves, while they can be identified by the analysis