

Reference excerpt 02-036

Source: arXiv:1604.05177v1 | License: CC BY 4.0 | Word window: 1329-1588

T of the ECG are indicated. The average and the standard deviation of these three parameters is calculated. The mean values and standard deviations of the parameters ta_R , tx_T and ta_P are presented in Table 2. Figure 3: Examples of CSA and ECG trace are shown. The waves a, x, v e y are correctly detected by an algorithm. The P, R e T events of the ECG trace are indicated. Figure 4: Examples of CSA and ECG trace are shown. The x and v e y are not correctly detected. Correct waves are indicated. The P, R e T events of the ECG are indicated.

3 DISCUSSION This note describes a methodology to extrapolate some interesting quantitative parameters from the IJV CSA trace, which is an analogue of the JVP. There are two important aspects that must be considered: i) the contouring of the IJV on each sonogram of the video-clip must be as accurate as possible. This means that the algorithm presented in [Sisini et al.(2015)] must be improved to achieve an accuracy close to 100 %, while the current one is about 95 %. ii) The wave detection algorithm must be improved. In fact it is not always sufficient to identify the waves, while they can be identified by the analysis of tracks performed by a human expert. This paper has not discussed the clinical use of these parameters.

ACKNOWLEDGEMENT The author wants to thank Giacomo Gadda, Valentina Tavoni and Valentina Sisini to have kindly reviewed the manuscript. REFERENCES